

Interactive comment on “An empirical method for absolute calibration of coccolith thickness” by Saúl González-Lemos et al.

Saúl González-Lemos et al.

heather.stoll@erdw.ethz.ch

Received and published: 23 November 2017

We have corrected this typographic error in the text. The microscope set up was indeed correct. Therefore, as you say, the other $\lambda/4$ plate should be below the upper linear polarizer, and not above it.

Interactive comment on Biogeosciences Discuss., <https://doi.org/10.5194/bg-2017-249>, 2017.